



Declaration of **CE** Conformity

I, Nigel Leigh, of Commtest Instruments Ltd., declare that the following equipment complies with EC standard IEC/EN 61326 (tested September 2004).

Vibration Analyzer - vbOnline

This equipment complies with the essential requirements of the EMC Directive 89/336EEC of the European Community.

Authorized Representative:

Nigel Leigh

A handwritten signature in blue ink that reads "Nigel Leigh". The signature is written over a horizontal line.

Date:

31-Oct-05

Parkside Laboratories



GLOBAL TESTING, CALIBRATION & CERTIFICATION SERVICES

◆ ELECTRICAL ◆ EMC ◆ ENVIRONMENTAL ◆

International Accreditation New Zealand (IANZ) has a Mutual Recognition Arrangement (MRA) with the National Association of Testing Authorities (NATA), Australia, such that both organisations recognise accreditations by IANZ and NATA as being equivalent. Users of test reports are recommended to accept test reports in the name of either accrediting body.

Laboratory Test Report

*This report may not be reproduced except in full.
It contains no corrections or erasures.*

Report No.: 8165-00

Date Issued: 9 September 2004

SUBJECT: VBOonline Vibration Data Collector
Model: Rev A
Serial Number: 20001

REQUESTED BY: Commtest Instruments Limited
28B Moorhouse Ave
P.O. Box 9297
Christchurch
NEW ZEALAND

INSTRUCTIONS: Test for compliance with BS EN 61326:1998 "Electrical equipment for measurement, control and laboratory use - EMC requirements".

CONTENTS: General
Test Specification
Date of test
Description
Results: BS EN 61326:1998
Appendix A: CISPR 16-1-1:2003
Appendix B: CISPR 16-1-2:2003
Appendix C: CISPR 16-1-3:2004
Appendix D: CISPR 16-1-4:2003
Appendix E: BS EN 61000-4-2:1995
Appendix F: BS EN 61000-4-3:2002
Appendix G: BS EN 61000-4-4:1995
Appendix H: BS EN 61000-4-5:1995
Appendix I: BS EN 61000-4-6:1996
Appendix J: BS EN 61000-4-8:1994
Appendix K: BS EN 61000-4-11:1994
Scans: - Radiated Emissions 30 MHz – 100 MHz
- Radiated Emissions 100 MHz – 1 GHz
- Conducted Emissions 150 kHz – 30 MHz

SUMMARY: All test results in this report in relation to the VBOonline Vibration Data Collector confirmed that the specimen Complied with the relevant provisions of BS EN 61326:1998.

APPROVED BY:

TESTED BY:

PREPARED BY:

SHEH ANWAR
IANZ Signatory

p.p. Brian Drumm
EMC Compliance Engineer

Clare Rochford
Technical Writer